

CAS 140D IR

The new reference array spectrometer for IR measurements

Key features at a glance

- ▲ Highly accurate infrared (IR) measurements
- ▲ High-performance InGaAs sensor with greatly improved optical bench design
- ▲ High signal sensitivity
- ▲ Low stray light
- ▲ Fast electronics



The new CAS 140D IR guarantees proven high measurement accuracy and reliability. It also promises significantly better performance for wavelengths in the infrared range than its predecessor CAS 140CT IR. Thanks to a new design of the optical architecture, throughput has been increased by up to 70 %. The stray light proportion has again been significantly reduced. In addition, the electronic platform of the CAS 140D IR enables faster processing of measurements: the minimum possible integration time has been reduced from 10 ms to 1 ms, the total scanning time for a measurement from 16 ms to 9 ms. The advantages

for the user lie in higher productivity due to shorter measuring times, accompanied by greater precision and repeatability. Like all Instrument Systems spectrometers, the infrared models feature calibration traceable to PTB or NIST. With a trigger box, various different CAS models can be combined to a MultiCAS system and complex spectral measurements run parallel over an extremely wide wavelength range. In the right combination with 2D imaging colorimeters, IR cameras, integrating spheres or goniometer systems they precisely and reliably perform individual customer tasks for wavelength ranges from UV to IR.

\\ TECHNICAL SPECIFICATIONS

Model	CAS 140D 171
Wavelength range	780 – 1700 nm
Detector	InGaAs
Pixel number	512
Spectral resolution	9 nm
Datapoint interval	2.1 nm
Wavelength accuracy	±0.5 nm
Integration time	1 ms - 3 s
Shortest duration SOT to EOT ¹⁾	6 ms
Sensor dynamic range	9000 : 1
Non-Linearity	±1 %
Cooling	TEC @ -10 °C

\\ TECHNICAL SPECIFICATIONS

Model	CAS 140D 171
Stray light	
Broadband for Illuminant A ²⁾	0.08 %
For LED ³⁾	0.03 %
With Laser ³⁾	0.015 %
Sensitivity	
Irradiance ⁴⁾	$1 \times 10^{-6} - 1.5 \times 10^3 \text{ W/m}^2\text{nm}$
Radiant flux ⁵⁾	$5 \times 10^{-6} - 4 \times 10^3 \text{ W}$
Spectrophotometry	
Baseline noise ⁶⁾	$\pm 0.05 \%$
Transmission measuring accuracy ⁷⁾	$\pm 0.2 \% \text{ T}$
Baseline drift ⁸⁾	0.1 counts/h
Spectrograph	
Focal length, grating	Approx. 120 mm f/3.5 / plane ruled grating
Slit	100 μm , 50 μm
Filter wheel / Shutter	Max. 7 slots for density filters OD 0.5 to OD 3.5; position monitoring with encoder
Electrical data	
AD converter	24 Bit resolution, Chip 16 Bit
PC interface	USB 2.0
Triggering	1 TTL input with ascending slope; 2 software-controlled TTL outputs; 1 TTL output with flash pulse
Other	
Dimensions (H, W, D)	144 x 341 x 359 mm ³
Power supply	Wide-range input 100 – 240 VAC 50/60 Hz
Power consumption	Max. 70 VA
Ambient temperature	15 – 35 °C; relative humidity 0 – 70 % max., non-condensing
Weight	Approx. 9 kg
Valid standards	In conformity with EN 60721-4-7 Class 7M2, EN 60721-4-7 Class 2M2, EN 61326:2004-05 and EN 61010-2002-08

Instrument Systems is continually working to develop and improve its products. Technical changes, errors or misprints do not constitute grounds for compensation. The company's terms of delivery and payment apply in all other respects.

¹⁾ With USB interface.

²⁾ Measured with long pass filter FELH1100 at 900 nm, relative to peak intensity of unweighted spectral data.

³⁾ Measured 150 nm left of peak wavelength, relative to peak intensity of unweighted spectral data.

⁴⁾ Measured with optical probe EOP-120 at 1200 nm and signal/noise ratio of 10:1, without averaging

⁵⁾ Measured with integrating sphere ISP 150L. Applies to a signal-to-noise ratio of 10:1.

⁶⁾ With shortest integration time, without averaging and with 50 % modulation. This value improves with appropriate averaging

⁷⁾ Applies to optimum spectral range; with 10 % transmission and immediately after recording an averaged baseline.

⁸⁾ Typical value. Applies for 1 hour warm-up.